

INDIAN INSTITUTE OF TECHNOLOGY JODHPUR
NH-62, Nagaur Road, Karwar, Jodhpur-342037
Email: sps@iitj.ac.in, Tel: 0291-280-1101

Corrigendum 3
Date: January 11, 2019

E-PROCUREMENT OF Rotary Microtome
NIT No: IITJ/SPS/2018-2019/51 dated 13 December 2018

Revised Technical Specification

SN	Technical specifications for Rotary Microtome	Technical Compliance (Yes/No)
1.	Complete unit with all accessories and one time consumables	
2.	Section thickness setting range: 0.5 – 60 μm	
3.	Horizontal specimen feed: 25 mm	
4.	Vertical specimen stroke: 59 mm	
5.	Specimen retraction: ON/OFF approx. 20 μm	
6.	Coarse feed wheel turn direction: User selectable	
7.	Specimen orientation: XY – $\pm 8^\circ$	
8.	Trimming thickness: 10 μm , 50 μm	
9.	Operating temperature range: 10°C to 35°C	
10.	Dimensions (W x D x H): 438 mm x 472 mm x 265 mm approx.	
11.	Warranty: 03 years	
12.	US-FDA certification	