

INDIAN INSTITUTE OF TECHNOLOGY JODHPUR

Old Residency Road, Ratanada, Jodhpur- 342 011

Corrigendum 1

Dated: January 1, 2015

Tender for Supply & Installation of Surface Area Analyzer

NIT No.: IITJ/SPS/ENERGY/2/1(I)/2014-15/58

ANNEXURE – I

Item: Surface Area Analyzer

- **Technical Specifications for Surface Area Analyzer:**

Point no.2. Analysis Gases: N₂, O₂, Ar, Kr, CO₂, CO, H₂, HCl, Br₂ and hydrocarbon

Point no.4. Operating Specifications:

Surface Area Range: Greater than 0.0003 m²/g

Pore Diameter Range: 2.0 - 5000 Angstroms

- “Last Date of submission: January 05, 2015 by 3:00 PM”

Now read as:

- **Technical Specifications for Surface Area Analyzer:**

Point no.2. Analysis Gases: N₂, O₂, Ar, Kr, CO₂, CO, H₂, and hydrocarbon

Point no.4. Operating Specifications:

Surface Area Range: Greater than 0.0005 m²/g

Pore Diameter Range: 3.5-5000 Angstroms

- “Last Date of submission: January 12, 2015 by 3:00 PM”

Bids will be opened on the same day at 03:15 PM.

Other Terms & Conditions will remain same.
