

# INDIAN INSTITUTE OF TECHNOLOGY JODHPUR

Old Residency Road, Ratanada, Jodhpur- 342 011

## Corrigendum 1

Date: May 09, 2015

Tender for Supply & Installation of Mass Flow Controller

NIT No.: IITJ/SPS/BIO/RD/1/1(I)/2015-16/5

Annexure-I

Mass Flow Controller; Qty.02 nos.

Technical Specifications:

- Accuracy :  $\pm$  (0.8% of reading + 0.2 % of full scale)
- Pre-Calibration : The device should be pre-calibrated with more than 25 gases and should be certified for such record.
- “Last Date of submission: May 12, 2015 by 3:00 PM”

## Now read as:

Annexure-I

Mass Flow Controller; Qty.02 nos.

Technical Specifications:

- Accuracy : Accuracy  $\pm$ 1% of the full scale
- Pre-Calibration : The MFC should be either precalibrated or easily calibratable for any pure gas or gas mixtures such as biogas, syngas etc.
- “Last Date of submission: May 18, 2015 by 3:00 PM”

Bids will be opened on the same day at 04:00 PM.

Other Terms & Conditions will remain same.

\*\*\*\*