

# INDIAN INSTITUTE OF TECHNOLOGY JODHPUR

Old Residency Road, Ratanada, Jodhpur- 342 011

## Corrigendum

Date: November 14, 2014

### **Tender for Supply & Installation of the Xenon-Arc Lamp**

**NIT No.: IITJ/SPS/ENERGY/2/1(I)/2014-15/54**

Technical specifications with necessary accessories:

- (A) Technical specification of the Lamp:

Gas filled : Xenon

Arc type : Short Arc Type

Wattage : 3000 watts

Voltage 24-34 volts

Current : 92 to 110 Amp.

Diameter of the lamp: 4 to 9 mm

Luminance : 120000 to 130000 lm

Diameter : 5 to 8 mm

length : 40 to 85 cm

Life-span of the Lamp : 1400-2200 Hours

- “Last Date of submission: November 25, 2014 by 3:00 PM”

## **Now read as:**

- (A) Technical specification of the Lamp:

Gas filled : Xenon

Arc type : Short Arc Type

Wattage : 3000 watts

Voltage 24-34 volts

Current : 92 to 110 Amp.

Diameter of the lamp: 4 to 9 mm

Luminance : 120000 to 130000 lm

Total length of the lamp: 40 to 85 cm

Life-span of the Lamp : 1400-2200 Hours

- “Last Date of submission: November 28, 2014 by 3:00 PM”

Bids will be opened on the same day at 04:00 PM.

Other Technical and Commercial Terms & Conditions will remain same.

\*\*\*\*